

No: NFSU/HQ/38(08)/2023-PUR/CPPP16/4208

Date: 08/11/2023

CORRIGENDUM

With reference to the tender advertisement which was published Tender notice **ref no: NFSU/HQ/38/2023-PUR/4027, 30th October, 2023** on CPPP Portal. The tender Id and tender reference no of above tender is (**ID: 2023_NFSU_778555_1, Ref No: NFSU (HQ)/38(08)/2023-PUR./CPPP16 published on 31st October, 2023**), and on NFSU's Website (i.e.www.nfsu.ac.in) by National Forensic Sciences University, the following addendum is issued.

Sr. No	Clause Reference	Existing Specifications	Revised Specifications
1.	Technical Specification (Technical Compliance Report) <ul style="list-style-type: none"> Page no : 18 Annexure – I Sr. No: 08 	"Specimen tilt: -10to80°orbetter,Rotation: 360°."	" Specimen tilt: -50to80°orbetter,Rotation: 360°."
2.	Technical Specification (Technical Compliance Report) <ul style="list-style-type: none"> Page no : 19 Annexure – I Sr. No: 14 	"19 inch LCD/TFT monitor, live images of different detector to be viewed simultaneously, full image live display, split live image, built in digital zoom, built in magnification, Pseudo color, Histogram display, Digital images should have resolution better than 2000X2000pixel sand should besaVedas BMP, TIFF or JPEG format."	"32-inch LCD/TFT monitor, live images of different detector to be viewed simultaneously, full image live display, split live image, built in digital zoom, built in magnification, Pseudo color, Histogram display, Digital images should have resolution as 10,000x10,000 pixel or better which is must for good quality images/poster prints."
3.	Technical Specification (Technical Compliance Report) <ul style="list-style-type: none"> Page no : 20 Annexure – I Sr. No: 19 	"Energy dispersive X-ray detector(EDS):EDS should have a minimumarea of 10mm ² . It should bepeltier cooling system."	"Energy dispersive X-ray detector(EDS):EDS should have a minimumarea of 30mm ² . It should bepeltier cooling system."

Gandhinagar Campus & Headquarter

Sector-9, Gandhinagar
Gujarat 382007
Ph:079-23977102/103
Fax: 079-23247465
Email:

Campus - director_gnr@nfsu.ac.in
HQ - exe_registrar@nfsu.ac.in

Delhi Campus

LNJN NICFS
Sector - 3
Outer Ring Road
Rohini, Delhi -110085
Ph:011-2752109, 27511580
Fax:011-27511571
Email: director_dc@nfsu.ac.in

Goa Campus

Curti, Ponda
Goa - 403401
Ph: 0832-2313036/3034
Email:director_goa@nfsu.ac.in

Tripura Campus

VIP Road, Radhanagar
Adjacent to Buddha Mandir
Agartala-799001, Tripura
Ph: 0381-2310009/0006,
2312525/2828
Email: director_tripura@nfsu.ac.in

Bhopal Campus

NFSU, C/o CFSL,
Barkhera Bonder,
P. O. Bairagarh Kalan,
Bhopal-462 030 (MP)
Ph:7552995271
Email: director_bhopal@nfsu.ac.in

Sr. No	Clause Reference	Additional Specification
1.	In addition specification Point add in Technical Compliance Report (Annexure-I)	“It should be a floor model scanning electron microscope (SEM), with all required accessories for basic operation”.

All other terms and conditions of the original tender document remain unchanged. Bidders are requested to take note of these corrections while preparing their bids.

This corrigendum shall form an integral part of the original tender document.



Sd/-
Executive Registrar
NFSU, Gandhinagar

Gandhinagar Campus & Headquarter

Sector-9, Gandhinagar
Gujarat 382007
Ph: 079-23977102/103
Fax: 079-23247465
Email:

Campus - director_gnr@nfsu.ac.in
HQ - exe_registrar@nfsu.ac.in

Delhi Campus

LNJN NICFS
Sector - 3
Outer Ring Road
Rohini, Delhi -110085
Ph: 011-2752109, 27511580
Fax: 011-27511571
Email: director_dc@nfsu.ac.in

Goa Campus

Curti, Ponda
Goa - 403401
Ph: 0832-2313036/3034
Email: director_goa@nfsu.ac.in

Tripura Campus

VIP Road, Radhanagar
Adjacent to Buddha Mandir
Agartala-799001, Tripura
Ph: 0381-2310009/0006,
2312525/2828
Email: director_tripura@nfsu.ac.in

Bhopal Campus

NFSU, C/o CFSL,
Barkhera Bonder,
P. O. Bairagarh Kalan,
Bhopal-462 030 (MP)
Ph: 7552995271
Email: director_bhopal@nfsu.ac.in