

F. No. G\_C\_/C.Pur/3-30/2024/1031

Date: -16.05.2025


**CORRIGENDUM NOTICE**

**Subject: Corrigendum — Revision of Technical Specifications for GeM Bid No: GEM/2025/B/6033497 (Procurement of HPLC under RUSA 2.0 for NFSU-Gandhinagar Campus)**

1. In reference to the GeM Bid No: **GEM/2025/B/6033497** for the procurement of **High-Performance Liquid Chromatography (HPLC)** under **RUSA 2.0**, it is informed to all concerned that the following amendment are made in technical specification;

Sl. No.	Earlier Specification	Revised Specification
1	<b><u>Technical specification Sl no : 2 ; point no 7</u></b> The flow accuracy must be $\pm$ 0.1% or better	<b><u>Technical specification Sl no : 2 ; point no 7</u></b> The flow accuracy must be $\pm$ 1% or better

2. All other terms and conditions of the Tender shall remain unchanged.  
3. All interested bidders are advised to carefully read the corrigendum before submitting their bids.  
4. The bidders are requested to take note of the above correction and submit the Technical bid accordingly.

  
**Campus Director**  
**NFSU, Gandhinagar**